



## Declaration of Conformity Electromagnetic Compatibility (CE mark)

Date of Issue: 4<sup>th</sup> September 2009  
Product Name: ChlordioXense

This product has been tested and found to be compliant with the following standards:

Standard	Title
EN 61000-6-1:2001	Electromagnetic compatibility (EMC) — Part 6-1: Generic standards — Immunity for residential, commercial and light industrial environments.
EN 61000-6-3:2001, A11:2004	Electromagnetic compatibility (EMC) — Part 6-3: Generic standards — Emission standard for residential, commercial and light industrial environments.

The ChlordioXense has been tested by the Compliancy Test House SGS under the name ChloroSense. A certificate from SGS for that instrument is available on request. The ChlordioXense instrument components and manufacture are the same as the ChloroSense instrument with the exception of the software calculations and therefore also passes EMC.

Signed on behalf of Palintest

A handwritten signature in red ink, appearing to read 'A. Ellis'.

Date 7 SEP 09

(Dr D. Ellis, Head of Compliance)

Palintest Ltd  
Registered Office:  
Palintest House, Kingsway,  
Tyne & Wear, England, NE11 0NS  
Tel: 0191 491 0808  
Fax: 0191 482 5372  
E-mail: [palintest@palintest.com](mailto:palintest@palintest.com)  
Web Site: [www.palintest.com](http://www.palintest.com)

A HALMA COMPANY

Registered in England Number: 1204118



VAT Registration Number: GB 178 8236 20