

30 MAR 2007

Palintest Ltd.  
Palintest House,  
Kingsway North,  
Team Valley Trading Estate,  
Gateshead. Tyne & Wear.  
NE11 0NS

Page 1 of 1

# Certificate of Test

## Electromagnetic Compatibility

### For

## Chlorosense

**Date of Issue:** 12<sup>th</sup> March 2007

**SGS Reference:** EMC107771/1/CL/07

**Details of Sample:** Model No.: Not Provided  
Unique Identifier: Not Provided

**Date of Receipt:** 9<sup>th</sup> March 2007

**Date of Test:** 9<sup>th</sup> March 2007 To 12<sup>th</sup> March 2007


**Test Required:** Test for compliance with: -  
EN 61000-6-1: 2001  
EN 61000-6-3: 2001, A11: 2004

**Conclusion:** In the configuration tested the sample meets the requirements of the standards detailed above.

**Issue Number:** 1

*This certificate is issued by the Company subject to its General Conditions of Service available upon request and accessible at [www.sgs.com](http://www.sgs.com). Attention is drawn to the limitations of liability, indemnification, jurisdictional, representative restrictions of test samples and sample retention policies defined therein. This certificate refers only to the sample submitted for test.*

*This certificate shall not be reproduced except in full without the written approval of SGS UK Ltd.*



*Authorised Signatory*

**A. Reynard / F. Huggins**  
**Technical Manager / Quality Manager**