

# Certificate of Test

Electrical Safety  
 for  
**ChloroSense**

**Date of Issue:** 11<sup>th</sup> April 2008

**SGS Reference:** ELS117891/1/C/MS/08

**Details of Sample:** Model: ChloroSense Serial No. P58, P67

**Date of Receipt:** 9<sup>th</sup> April 2008

**Date of Test:** 9<sup>th</sup> April 2008 to 11<sup>th</sup> April 2008


**Test Specification:** Test for compliance with BS EN 60529:1992 AMD1 7643:1993  
 AMD2 10931:2000 (Degrees of protection provided by enclosures  
 (IP code) product tested to IP67

**Conclusion:** The sample tested was found to comply with the Test Specification.

This document is issued by the Company under its General Conditions of Service accessible at [http://www.sgs.com/terms\\_and\\_conditions.htm](http://www.sgs.com/terms_and_conditions.htm). Attention is drawn to the limitation of liability, indemnification and jurisdiction issues defined therein. Any holder of this document is advised that information contained hereon reflects the Company's findings at the time of its intervention only and within the limits of Client's instructions, if any. The Company's sole responsibility is to its Client and this document does not exonerate parties to a transaction from exercising all their rights and obligations under the transaction documents. Any unauthorized alteration, forgery or falsification of the content or appearance of this document is unlawful and offenders may be prosecuted to the fullest extent of the law."

Unless otherwise stated the results shown in this test report refer only to the sample(s) tested and such sample(s) are retained for 28 days only. This document cannot be reproduced except in full, without prior approval of the Company.  
 Tests marked(\*) are not UKAS accredited

**Test Engineer: M. Stout**



**Authorised Signatory: R.N. Gough**



All enquires relating to this certificate must be directed to the Authorised Signatory

QUAL 08 Rev