

Palintest Ltd
Palintest House
Kingsway North
Team Valley Trading Estate
Gateshead
NE11 0NS

Page 1 of 1

Certificate of Test

Electrical Safety
for
Photometer

Date of Issue: 16th October 2009

SGS Reference: ELS129278/2/C/MG/09

Details of Sample: Model: 'Compact' photometer range
Rated: 2 x 1.5V AA internal battery
S/No: (IPX7) 1009808101, (IP6X) No serial number; ELS sample 1 of 2

Date of Receipt: 6th October 2009

Date of Test: 8th October 2009 to 16th October 2009

Test Specification: Test for compliance with IP67 as described in BS EN 60529:1992 incorporating Amendments 1 & 2, Degrees of protection provided by enclosures (IP code)

Conclusion: The sample tested was found to comply with the Test Specification.

Test Engineer: M.W. Garfitt

Counter Signatory: S. Wilkinson



All enquires relating to this certificate must be directed to the Test Engineer

This document is issued by the Company subject to its General Conditions of Service available on request or accessible at www.sgs.com/terms_and_conditions.htm and, for electronic format documents, subject to Terms and Conditions for Electronic Documents at www.sgs.com/terms_e-document.htm. Attention is drawn to the limitation of liability, indemnification and jurisdiction issues defined therein. Any holder of this document is advised that information contained hereon reflects the Company's findings at the time of its intervention only and within the limits of Client's instructions, if any. The Company's sole responsibility is to its Client and this document does not exonerate parties to a transaction from exercising all their rights and obligations under the transaction documents. This document cannot be reproduced except in full, without prior written approval of the Company. Any unauthorized alteration, forgery or falsification of the content or appearance of this document is unlawful and offenders may be prosecuted to the fullest extent of the law." Unless otherwise stated the results shown in this test report refer only to the sample(s) tested and such sample(s) are retained for 28 days only.