

Palintest Ltd.
Palintest House, Kingsway North,
Team Valley Trading Estate.
Gateshead. Tyne & Wear.
NE11 0NS

Page 1 of 1

Certificate of Test

Electromagnetic Compatibility

For

Photometer

Date of Issue: 28th November 2007

SGS Reference: EMC114330/1/CL/07

Details of Sample: Model No.: 7500
Unique Identifier: 1022-ff47

Date of Receipt: 28th November 2007

Date of Test: 28th November 2007


Test Required: Test for compliance with: -
EN 61000-6-1: 2001
EN 61000-6-3: 2001, A11: 2004

Conclusion: In the configuration tested the sample meets the requirements of the standards detailed above.

Issue Number: 1

This certificate is issued by the Company subject to its General Conditions of Service available upon request and accessible at www.sgs.com. Attention is drawn to the limitations of liability, indemnification, jurisdictional, representative restrictions of test samples and sample retention policies defined therein. This certificate refers only to the sample submitted for test.

This certificate shall not be reproduced except in full without the written approval of SGS UK Ltd.



Authorised Signatory

A. Reynard / F. Huggins
Technical Manager / Quality Manager